


4/17/2013

SIEVE SERIAL NUMBER

99200958



**DUAL**  
Manufacturing Co. Inc.

**CERTIFICATE OF  
INSPECTION**

**CERTIFICATE OF COMPLIANCE  
TO SPECIFICATIONS OF**

**ASTM- AMERICAN SOCIETY FOR TESTING Materials  
ANSI- AMERICAN NATIONAL STANDARDS INSTITUTE  
ISO- INTERNATIONAL STANDARDS INSTITUTE**

This is to certify that the openings in the wire cloth of this test sieve have been checked through advanced optical technology to assure conformity to ASTM specifications E11-09.

**Sieve specifications:**

ASTM mesh number - 40  
Mesh openings (inches) - 0.0165 inches  
Mesh openings (m.m.) - 0.425 mm

Wire diameter (inches) - 0.011 inches  
Wire diameter (m.m.) - 0.280 mm

**TEST RESULTS  
(per 60 openings checked)**

Mesh opening warp direction of mesh (inches) = 0.01640  
Mesh opening shute direction of mesh (inches) = 0.01685

Wire diameter (inches) = 0.01102

**SIEVE PASSED INSPECTION**

SIGNATURE \_\_\_\_\_

3522 Martens Street , Franklin Park, Illinois 60131  
Customer service 847-260-5370 Fax 847-260-5419 e-mail info@dualmfg.com

